

IN THE UNITED STATES PATENT AND TRADEMARK OFFICE
BEFORE THE BOARD OF PATENT APPEALS AND INTERFERENCES

Re: Application of: Werner KNEBEL et al.
Serial No.: 10/713,887 Confirmation No.: 1103
Filed: November 14, 2003
For: SCANNING MICROSCOPE
Art Unit: 2872
Examiner: Joshua L. Pritchett
Customer No.: 23280

Mail Stop: APPEAL BRIEF - PATENTS
Commissioner for Patents
P.O. Box 1450
Alexandria, VA 22313-1450

December 19, 2006

APPELLANTS' SUPPLEMENTAL REPLY BRIEF UNDER 37 C.F.R. § 41.41 IN RESPONSE
TO EXAMINER'S SUPPLEMENTAL ANSWER

Sir:

Appellants submit this reply brief for the consideration of the Board of Patent Appeals and Interferences (the "Board") in response to the Examiner's Supplemental Answer mailed October 19, 2006.

Independent claim 12 recites a method for scanning a sample (e.g., 25 in Fig. 1; e.g., specification lines 8-10), comprising: providing a microscope having a detector (e.g., 29 or 33 in Fig. 1; e.g., specification page 6, lines 12-14 or lines 17-21, respectively) disposed in a detection beam path (e.g., 27 in Fig. 1; e.g., specification page 6, lines 10-16) and configured to receive detection light (e.g., 27 in Fig. 1; e.g., specification page 6, lines 10-16) proceeding from the sample; measuring the light power level of the detection light using a monitoring means (e.g., 43 in Fig. 1; e.g., specification page 6, line 27 – page 7, line 1); and blocking the detection beam path, when the light power level of the detection light exceeds a definable threshold (e.g., specification page 7, lines 1-3), using an optical shutter means (e.g., 35 or 37 in Fig. 1; e.g., specification page 6, lines 23-25) disposed between the sample and the detector.

Whether claims 1, 3-12 and 14-20 should be rejected under 35 U.S.C. §102(b) as being anticipated by U.S. Patent No. 5,065,008 issued to Hakamata et al. ("Hakamata et al."). Whether claims 2 and 13 should be rejected under 35 U.S.C. 103(a) as being unpatentable over Hakamata et al. in view of U.S. Patent No. 5,523,573 issued to Hänninen et al. ("Hänninen et al.").

Rejection under 35 U.S.C. §102(b) to claims 1, 3-12 and 14-20 based on Hakamata et al:

{W:\20793\0204530us0\00942201.DOC}

of the detection light” and “blocking the detection beam path, when the light power level of the detection light exceeds a definable threshold, using an optical shutter means.”

It is respectfully submitted that the anticipation rejection represents clear error because the cited prior art reference, Hakamata et al., fails to describe, expressly or inherently, every element as set forth in the claims. See MPEP 2131. Specifically, the cited prior art reference fails to teach the above-recited limitations of claims 1 and 12. Hakamata et al. does not control liquid crystal shutters 23a “based on the light power level of the detection light exceeding a definable threshold,” as recited in claim 1, or use the shutters 23a for “blocking the detection beam path, when the light power level of the detection light exceeds a definable threshold,” as recited in claim 12. Hakamata et al. does not control the shutters based on a light power level at all. In contrast Hakamata et al. explicitly teaches controlling the shutters in synchronization with a scanning operation based on a synchronization signal received from a control circuit 32. See Hakamata et al. col. 8, lines 29-39, and Fig. 1. Hakamata et al. describes a photodetector 25 having a receiving surface 25a upon which a laser beam 11’ impinges after passing “through a liquid crystal shutter 23a, which has been set to the open state.” See Hakamata et al., col. 7, lines 58-63. There is no disclosure in Hakamata et al. that any output, such as a power level of detected light, of photodetector 25 affects the liquid crystal shutters 23a in any way. Fig. 1 of Hakamata et al., which shows no signal output from photodetector 25 into control circuit 32, provides further substantiation that the shutters 23a are not operated based on the light power level of the detection light, as required by claims 1 and 12. Because Hakamata et al. fails to teach the above limitations of independent claims 1 and 12, anticipation of these claims, or any of their dependent claims, has not been established.

Withdrawal of the rejection of claims 1, 3-12 and 14-20 under 35 U.S.C. §102(b) as being anticipated by Hakamata et al. is respectfully requested.

Rejection under 35 U.S.C. §103(a) to claims 2 and 13 based on Hänninen et al:

Hänninen et al. does not teach or suggest the features of claims 1 and 12 missing from Hakamata et al. (see argument above relative to the rejection under 35 U.S.C. §102(b) based on

Hakamata et al.). Hänninen et al. merely describes a shutter S1 used to chop light in time-resolved methodology. See col. 8, lines 15-18, and Fig. 5. Therefore a combination of Hakamata et al. and Hänninen et al. could not teach or suggest all the features recited the dependent claims 2 and 13.

Withdrawal of the rejection of claims 2 and 13 under 35 U.S.C. §103(a) as obvious over Hakamata et al. in view of Hänninen et al. is respectfully requested.

8. RESPONSE TO EXAMINER'S SUPPLEMENTAL ANSWER MAILED OCTOBER 19, 2006

The claims require operating the optical shutters based on, i.e., in dependence upon, a measured light power level. The Hakamata et al. reference in no way teaches or suggests this. Hakamata clearly and explicitly controls the liquid crystal shutter 23a based on a synchronization from control circuit 32, which receives no measured power level input. See Hakamata et al., col. 8, lines 29-39, and Fig. 1. Whatever effect the liquid crystal shutter 23a may have on the detected light at photodetector 25, the operation of shutter 23a in Hakamata et al. is in no way based on such detected light.

It is respectfully submitted that the claimed invention is patentable over Hakamata et al.

APPENDIX A

PENDING CLAIMS 1-5 and 8-20 OF
U.S. APPLICATION SERIAL NO. 10/713,887

Claim 1 (previously presented): A scanning microscope having a detector, arranged in a detection beam path, for receiving detection light proceeding from a sample, a monitoring means that measures the light power level of the detection light, and an optical shutter means between the sample and the detector with which the detection beam path can be blocked based on the light power level of the detection light exceeding a definable threshold.

Claim 2 (original): The scanning microscope as defined in Claim 1, wherein the detector is a non-descan detector.

Claim 3 (original): The scanning microscope as defined in Claim 1, wherein the detector is a descan detector.

Claim 4 (original): The scanning microscope as defined in Claim 1, further comprising a control means for controlling the shutter means.

Claim 5 (original): The scanning microscope as defined in Claim 1, wherein the detection beam path can be automatically opened up before the beginning of a scanning operation, and blocked at the end of the scanning operation.

Claims 6-7 (canceled)

Claim 8 (original): The scanning microscope as defined in Claim 4, wherein the control means extrapolates the future change over time in the detection light power level, and wherein the

Claim 15 (previously presented): The method as defined in Claim 12, further comprising controlling the shutter means using a control means.

Claim 16 (previously presented): The method as defined in Claim 12, further comprising automatically opening the detection beam path before the beginning of a scanning operation, and blocking the detection beam path at the end of the scanning operation.

Claim 17 (previously presented): The method as defined in Claim 15, further comprising extrapolating the future change over time in the detection light power level using the control means.

Claim 18 (previously presented): The method as defined in Claim 12, wherein the shutter means contains a mechanical shutter or an electrooptical element or acoustooptical element or LCD element.

Claim 19 previously presented): The method as defined in Claim 12, wherein the detector contains a photodiode, in particular an avalanche photodiode, or a CCD element, in particular an EMCCD element, or a photomultiplier or photomultiplier array.

Claim 20 (previously presented): The method as defined in Claim 12, wherein the scanning microscope is a confocal scanning microscope.

APPENDIX B

Evidence Appendix under 37 C.F.R. §41.37 (c) (ix):

No evidence pursuant to 37 C.F.R. §§1.130, 1.131 or 1.132 and relied upon in the appeal has been submitted by appellants or entered by the examiner.

APPENDIX C

Related proceedings appendix under 37 C.F.R. §41.37 (c) (x):

As stated in “2. RELATED APPEALS AND INTERFERENCES” of this appeal brief, appellants, their legal representatives, and assignee are not aware of any appeal or interference that directly affects, will be directly affected by, or will have a bearing on the Board’s decision in this appeal.